

# Quality & Calibration Report



131 Business Center Drive, A-3  
Ormond Beach, FL 32174

TEL: [386.256.4910](tel:386.256.4910) | [386.310.4933](tel:386.310.4933) M-F 9-5pm EST

FAX: 866.422.2356

EMAIL: [support@co2meter.com](mailto:support@co2meter.com) | [www.co2meter.com](http://www.co2meter.com)

## Certificate of Calibration and Testing

Date: 8/13/2021  
Order Number: 19884 RMA#0542-r3e-19  
Customer: Climate Change Truth Inc.

Item	Qty	Serial Number	Calibration Gas	Reading	Data-Logging Test
CM-501			Nitrogen 100% (±0.015%)	0 ppm(± 10 ppm)	Pass
CO2 PORTABLE	2	E9113000035	P/N NIUHPP55		
DIFFUSION		E9113000038	LOT#304-402053940-1		
CO2/Temp			CO2 400 PPM (± 50 PPM)	400 ppm ( ± 50 ppm )	Pass
/Rh/Baro			P/N 221L-34-400		
			LOT# IBH-34-400-1		

QA Inspector



We are committed to providing a quality product.  
This instruments has undergone rigorous testing throughout its manufacture.  
This instrument has been factory Inspected , Tested & Calibrated  
in accordance with the conditions and requirements of our Quality System,  
Operating Standards & Sales Agreements.  
All calibration gases are traceable to industries standards.

**\*\*NOTE \*\* This certification is valid for 180 days from date issued.**

**Calibration of this device is traceable to NIST standards**